## CAS DEVICE UPDATE

| Company<br>Name  | Stent Name  | Tapered Stents                        |                 |  |  | Status                                  | Embolic Protection Devices                                     |  |          | Status  |
|--|---|---------------------------------------|-----------------|--|--|---|--|--|----------|---|
|  |   | Diameter<br>(Proximal/Distal)<br>(mm) | Lengths<br>(mm) | Diameters<br>(mm)                      | Lengths<br>(mm)                                      | Within<br>Carotid<br>Market             | Name   | Type                                     | Position | Within<br>Carotid<br>Market                   |
| Abbott<br>Vascular   | RX Acculink<br>Carotid Stent<br>System<br>Xact Carotid<br>Stent | 10/7, 8/6<br>10/8, 9/7, 8/6           | 30, 40          | 5, 6, 7, 8, 9,<br>10<br>7, 8, 9, 10    | 20, 30, 40   | Approved<br>in and<br>outside<br>the US | RX Accunet  EmboShield  Nav <sup>6</sup>                       | Filter Filter BareWire delivery          | Distal   | Approved<br>in and<br>outside<br>the US       |
| Boston<br>Scientific<br>Corporation                        | Carotid<br>Wallstent<br>Endoprosthesis                          | N/A                                   | N/A             | 6, 8, 10<br>(uncon-<br>strained)       | 21, 22,<br>24, 29,<br>36, 37<br>(uncon-<br>strained) | Approved<br>in and<br>outside<br>the US | FilterWire<br>EZ Embolic<br>Protection<br>System               | Filter                                   | Distal   | Approved<br>in and<br>outside<br>the US       |
| Cordis<br>Corporation                                      | Precise Nitinol<br>Stent<br>Precise Pro RX                      | Autotapering                          | 20, 30, 40      | 5, 6, 7, 8, 9,<br>10<br>5, 6, 7, 8, 9, | 20, 30, 40   | Approved<br>in and<br>outside<br>the US | Angioguard<br>XP Emboli<br>Capture<br>Guidewire<br>Angioguard  | Filter                                   | Distal   | Approved<br>in and<br>outside<br>the US       |
|  | Nitinol Stent   |                                       |                 | 10                                     |  |   | RX Emboli<br>Capture<br>Guidewire                              |  |          |   |
| ev3 Inc.   | Protégé RX  | 10/7, 8/6                             | 30, 40          | 6, 7, 8, 9, 10                         | 20, 30 40,<br>60                                     | Approved in and outside the US          | SpiderFX   | Filter-<br>preferred<br>wire<br>delivery | Distal   | Approved<br>in and<br>outside<br>the US       |
| Invatec, Inc.  | Embolic protec  | tion device only                      |                 |  |  |   | Mo.Ma<br>Ultra<br>Proximal<br>Cerebral<br>Protection<br>Device | Proximal<br>occlusion                    |          | Approved<br>in and<br>outside<br>the US       |
| Lumen<br>Biomedical,<br>distributed<br>by Invatec,<br>Inc. | Embolic protection device only                                  |                                       |                 |  |  |   | FiberNet<br>Embolic<br>Protection<br>System                    | Filter                                   | Distal   | Approved<br>in and<br>outside<br>the US       |
| W. L. Gore &<br>Associates                                 | Embolic protec  | tion device only                      |                 |  |  |   | Gore Flow<br>Reversal<br>System                                | ICA flow<br>reversal                     | Proximal | Cleared in<br>the US; CE<br>Mark in<br>Europe |