

STENTS

SELF-EXPANDING STENT GRAFTS AVAILABLE FOR SALE IN THE US

Company Name	Product Name	Indicated Use	Material Used	Introducer Size (F)	Endhole	Stent Diameter (mm)	Stent Length (mm)	Delivery System Length (cm)
Bard Peripheral Vascular, Inc.	Flair	Stenosis at venous anastomosis of ePTFE or other synthetic AV grafts	Nitinol/ePTFE	9	0.035	6, 7, 8, 9	30, 40, 50	80
LeMaitre Vascular, Inc.	aSpire	Tracheobronchial malignant neoplasms	ePTFE covered nitinol	7	0.018	6, 7, 8, 9	20, 30, 50, 100	50
W. L. Gore & Associates	Gore Viabahn Endoprosthesis, Gore Viabahn Endoprosthesis With Heparin Bioactive Surface	5–8 indicated for SFA; 5–13 indicated for iliac artery; 9–13 indicated for tracheobronchial malignant neoplasms	Nitinol/ePTFE	7–12	0.035–0.025	5, 6, 7, 8, 9, 10, 11, 13	25, 50, 100, 150	75, 110, 120
	Gore Viabil Endoprosthesis	Malignant biliary strictures		10 percutaneous, 8.5 endoscopic	0.035	8, 10	40, 60, 80, 100	40 percutaneous, 200 endoscopic
	Gore Viatorr Endoprosthesis	TIPS, tracheobronchial		10	0.038	8, 10, 12	40, 50, 60, 70, 80	75

AV, arteriovenous; ePTFE, expandable polytetrafluoroethylene; SFA, superficial femoral artery; TIPS, transjugular intrahepatic portosystemic shunt.