

FDA Perspectives on Aortic Reinterventions

What is the gold standard for follow-up, and what makes good follow-up practices?

With Rohini Retarekar, PhD; Robert Lee, MD; Ronald Fairman, MD;
and Carmen Gacchina Johnson, PhD

One question we are exploring in this edition of *Endovascular Today* is the shift in the discussion surrounding reintervention after abdominal aortic aneurysm (AAA) repair given that recent data suggest reinterventions improve survival long term and may just be a marker of adequate follow-up. What is the FDA's current view on reinterventions? Does the agency consider the need for reintervention a failure, or does it largely imply that higher follow-up compliance equating to a larger number of reinterventions could be due to factors that go beyond device effectiveness failure? Are there any notable changes in the Agency's thinking regarding reinterventions?

Aortic disease is complex and dynamic, and it warrants lifelong surveillance regardless of the treatment method. Longer-term follow-up of aortic endovascular stent grafts shows that reinterventions continue to occur over time, and oftentimes, they are deemed necessary by the treating physician to ensure that the benefit of the medical device is not lost.¹ Reinterventions performed after AAA repair are one consideration of many that the Agency evaluates when assessing the benefits and risks of a given device. Of particular interest to FDA is the reintervention reason and type.² The overall rate of reinterventions and their outcomes can help inform the overall benefit/risk profile of the device.

In all FDA premarket approval (PMA) application decisions, the benefits and risks of the device are carefully considered,³ including information pertaining to reintervention. For instance, for novel technology that fulfills an unmet need, an elevated rate of reinterventions may be considered an acceptable risk if the benefits of using the device are profound (eg, reduction in morbidity and mortality compared to other available treatment options, device meets an unmet need). For other scenarios where

in there may be several established treatment options available that do not require reinterventions and a new device does not provide notable benefit, there may be less tolerance for accepting the risk of reinterventions. In making this benefit/risk evaluation, the reason and timepoint for reintervention are also important. For example, reinterventions performed to address device integrity events or failures of the device to function as intended are viewed differently than reinterventions associated with disease progression. Other factors also come into play, such as the nature of reinterventions. A treatment approach wherein the majority of reinterventions are outpatient procedures using minimally invasive techniques with good outcomes would be viewed more favorably by all stakeholders (physicians, patients, and FDA) as compared with those that are associated with increased morbidity and mortality risks and need inpatient care. The outcome of reinterventions is also critical (eg, successful resolution of issue, acknowledgment if the problem remains unresolved after multiple reintervention attempts, morbidity and mortality associated with the reinterventions). It is also important to consider patient perspectives, such as desire for minimally invasive treatment options, need for lifelong follow-up, and awareness of the potential need for reinterventions. A key part of FDA's role in reviewing PMAs for aortic devices is ensuring that physicians and patients are provided clear device-specific safety and effectiveness data to inform them of both expected outcomes and need for reinterventions (eg, via patient labeling and instructions for use [IFU]).

What are FDA's views on the importance of clearly defining reinterventions, the need for reintervention, and reintervention rates?

Use of standardized definitions for reinterventions and clear documentation of the reason, timepoint, type, frequency, and outcome of each reintervention is helpful

to all stakeholders. For example, it can allow FDA to evaluate outcomes across different treatment options while assessing the benefit/risk profile of a device, it can allow physicians to make informed treatment decisions for their patients, and it can allow patients to have a balanced understanding of the device-specific risks and benefits. FDA acknowledges that the Society for Vascular Surgery (SVS) publications on reporting standards for endovascular aneurysm repair (EVAR),⁴ thoracic endovascular aortic repair,⁵ and complex EVAR involving the renal and mesenteric vessels⁶ have helped lead to a more uniform adoption of definitions for reporting.

From a regulatory perspective, what is the current gold standard for follow-up after EVAR? What are some of the essential elements?

Pivotal study designs intended to support regulatory approval typically include 5-year follow-up of study patients with imaging at appropriate intervals (eg, CT scans at 30 days, 6 months, 1 year, and annually thereafter through 5 years). To support PMA approval of endovascular aortic stent grafts, complete 1-year follow-up with adequate imaging to evaluate all critical parameters is generally provided. If the 1-year data support a favorable benefit/risk profile of the device, 5-year follow-up data on pivotal study patients has been collected postmarket, and the results are shared in annual clinical updates and an IFU update once the study is complete.⁷

When the postmarket surveillance of EVAR devices was discussed at the 2021 Circulatory System Devices Panel meeting, the panel members recommended postmarket follow-up through 10 years.² The recommendation has been adopted by the medical device industry, as noted through new postapproval enrollment studies of recently approved endovascular and hybrid stent grafts,⁸⁻¹⁰ which are intended to address questions about the real-world performance of novel medical devices or when there are specific questions about the performance of approved devices in the longer term.

FDA continues to emphasize to physicians and other health care providers the critical importance of lifelong surveillance, including imaging, for patients with EVAR.¹¹ In real-world use of EVAR devices, physicians may find device-specific recommendations (eg, IFU recommendations, FDA safety communications) and the SVS clinical practice guidelines helpful for follow-up of patients.

What efforts might result in improved patient compliance in follow-up?

Maintaining patient follow-up compliance is a known challenge especially evident in postmarket studies and real-world surveillance of medical devices.⁷ As discussed

at the 2021 panel meeting,² improved follow-up compliance for EVAR starts with appropriate patient selection and clear communication from physicians that regular clinical follow-up and imaging are critical if EVAR is performed. This essential element is typically reflected in patient labeling and would benefit if clearly reinforced as part of the informed consent process.

Successful long-term follow-up has been achieved by committed device manufacturers and physicians working together, engaged study coordinators, and use of optimized follow-up strategies (eg, telemedicine, remote imaging). For example, data published on mid- and long-term outcomes from some manufacturer-sponsored investigational device exemptions (IDEs)¹² and postapproval studies¹³ show that high rates of follow-up compliance can be maintained in the longer term (eg, 5 years) when there are focused efforts to achieve this goal.

FDA is not committed to any one approach in terms of maximizing postmarket follow-up compliance and is ready to engage with all stakeholders regarding strategies to improve follow-up and data collection over the longer term.

What are important considerations for regulatory use of clinical data?

When clinical data are used for regulatory decision-making, the relevance and reliability of that data are considered carefully by FDA. As it relates to real-world evidence, the FDA guidance document “Use of Real-World Evidence to Support Regulatory Decision-Making for Medical Devices”¹⁴ outlines important considerations. Data quality control is essential for providing confidence in the reliability of data, and there are several important factors outlined in the guidance to consider, including standardizing procedures to ensure the use of uniform and systematic methods for collecting and cleaning data.

At the 2021 panel meeting, potential plans for long-term EVAR postmarket surveillance studies were discussed. Are there any updates regarding these studies and what they will comprise?

Since the panel meeting, FDA has participated in collaborative efforts led by industry, with engagement from multidisciplinary professional societies, and representatives from existing EVAR data collection infrastructure (eg, SVS Vascular Quality Initiative and Kaiser Permanente) regarding the development/strengthening of infrastructure that would allow for robust postmarket surveillance of EVAR devices.^{15,16} This collaboration has aligned on specific elements that are critical for this effort. These include collection of device-specific clinical and imaging (site-reported)

outcomes of interest through 10 years with an ability for further assessment on an as-needed basis (eg, core laboratory evaluation of imaging in case of a safety signal). Data collection will focus on including a diverse patient population (particularly underrepresented groups), various institution types (eg, urban, rural, academic, private), a wide geographic distribution, disparities in health care utilization, and physicians with a range of EVAR experience.^{16,17} FDA is enthusiastic about reaching alignment on the protocol and looks forward to continued collaboration as the improved surveillance is implemented, incorporating learnings from prior surveillance efforts.

How might the extensive sponsor-investigator experience in the endovascular graft space be used to shed light on the issues around reinterventions and patient follow-up? How might new models in real-world evidence collection be incorporated?

There have been several published reports from sponsor-investigator IDEs, particularly those evaluating endovascular devices for treatment of thoracoabdominal aortic disease.¹⁷ The follow-up compliance and treatment outcomes in these IDEs have been exceptional in many cases.¹⁸ Sustaining the longer-term treatment benefit often involves reinterventions. With complex multi-branch devices, the potential for device or treatment failure increases, making the need for rigorous follow-up of patients and indicated reinterventions even more critical. The FDA remains committed to working with all stakeholders to translate the enormous learnings from the sponsor-investigator IDE experience to safe and effective commercially available devices for all patients with complex aortic pathology in the United States.

FDA and other stakeholders are interested in obtaining real-world long-term outcomes data. However, it is acknowledged that intensive resources are involved in this type of postmarket data collection, particularly in maintaining a high level of follow-up compliance and collection of imaging data. Therefore, FDA is open to targeted strategies to engage select, yet diverse, sites (eg, in terms of geography, patient demographics, physician experience) committed to the long-term data collection for certain endpoints that may otherwise be challenging to obtain (eg, imaging-based outcomes).

Does FDA have any patient engagement data that provide perspectives on potential need for reintervention versus feelings regarding open surgery?

When making a benefit/risk determination, FDA recognizes that patient perspectives on benefits and risks may

include tolerance of risk to achieve a probable benefit, especially if that benefit results in an improvement in quality of life. FDA acknowledges that patients with aortic pathology may desire less invasive treatment options, even when that means more frequent follow-up and greater need for reinterventions as compared with open repair.¹⁹

FDA encourages sponsors to include patient preference information in marketing submissions, as noted in the 2016 guidance document, "Patient Preference Information – Voluntary Submission, Review in Premarket Approval Applications, Humanitarian Device Exemption Applications, and De Novo Requests, and Inclusion in Decision Summaries and Device Labeling."²⁰ FDA also works with industry to develop transparent documents (eg, IFU, patient labeling, Summary of Safety and Effectiveness Data) so that the risks and benefits of a device are clear and can be carefully considered for individual patient circumstances. It remains key that the informed consent process incorporates clear communication from physicians regarding the importance of lifelong surveillance with appropriate imaging and potential need for long-term reinterventions as their patients decide between endovascular or open aortic repair. ■

1. Columbo JA, Kang R, Hoel AW, et al. A comparison of reintervention rates after endovascular aneurysm repair between the Vascular Quality Initiative registry, Medicare claims, and chart review. *J Vasc Surg.* 2019;69:74-79.e6. doi: 10.1016/j.jvs.2018.03.423
2. US Food and Drug Administration. 24 hour summary of the Circulatory System Devices panel meeting: November 3, 2021. Accessed February 14, 2024. <https://www.fda.gov/media/154009/download>
3. US Food and Drug Administration. Factors to consider when making benefit-risk determinations in medical device premarket approval and de novo classifications. Guidance for industry and Food and Drug Administration staff. Accessed February 14, 2024. <https://www.fda.gov/media/99769/download>
4. Chaikof EL, Blankensteijn JD, Harris PL, et al. Reporting standards for endovascular aortic aneurysm repair. *J Vasc Surg.* 2002;35:1048-1060. doi: 10.1067/mva.2002.123763
5. Fillingim MF, Greenberg RK, McKinsey JF, et al. Reporting standards for thoracic endovascular aortic repair (TEVAR). *J Vasc Surg.* 2010;52:1022-1033.e15. doi: 10.1016/j.jvs.2010.07.008
6. Oderich GS, Forbes TL, Chaer R, et al. Reporting standards for endovascular aortic repair of aneurysms involving the renal-mesenteric arteries. *J Vasc Surg.* 2021;73:45-52S. doi: 10.1016/j.jvs.2020.06.011
7. US Food and Drug Administration. FDA executive summary: Circulatory System Devices panel meeting: November 3, 2021. General issues panel: real world surveillance of AAA endovascular stent grafts. Accessed February 14, 2024. <https://www.fda.gov/media/153647/download>
8. US Food and Drug Administration. Summary of safety and effectiveness data: Gore TAG thoracic branch endoprosthesis (P210032). Accessed February 14, 2024. https://www.accessdata.fda.gov/cdrh_docs/pdf21/P210032B.pdf
9. US Food and Drug Administration. Summary of safety and effectiveness data: Thoraflex hybrid device (P210006). Accessed February 14, 2024. https://www.accessdata.fda.gov/cdrh_docs/pdf21/P210006B.pdf
10. US Food and Drug Administration. Summary of safety and effectiveness data: RelayPro thoracic stent graft system (P200045). Accessed February 14, 2024. https://www.accessdata.fda.gov/cdrh_docs/pdf20/P200045B.pdf
11. US Food and Drug Administration. FDA Advisory Panel recommendations on lifelong surveillance and long-term postmarket data collection for patients with AAA endovascular aortic repair—letter to health care providers. Published February 28, 2022. Accessed February 14, 2024. <https://www.fda.gov/medical-devices/letters-health-care-providers/fda-advisory-panel-recommendations-lifelong-surveillance-and-long-term-postmarket-data-collection>
12. Oderich GS, Farber MA, Schneider D, et al. Final 5-year results of the United States Zenith Fenestrated prospective multicenter study for juxtarenal abdominal aortic aneurysms. *J Vasc Surg.* 2021;73:1128-1138.e2. doi: 10.1016/j.jvs.2020.08.128
13. US Food and Drug Administration. Post-approval studies database: Ovation Long Term Follow-up Study. Accessed February 14, 2024. https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmas/pma_csm.cfm?_id=489523&c_id=770
14. US Food and Drug Administration. Use of real-world evidence to support regulatory decision-making for medical devices. Guidance for industry and Food and Drug Administration staff. Accessed February 14, 2024. <https://www.fda.gov/media/99447/download>
15. Fairman R. Regulatory Viewpoint on Abdominal Aortic Aneurysm and Carotid Disease Trials. Presented at: 20th Global CardioVascular Clinical Trialists (CVCT) forum; December 1-2, 2023; Washington, DC.
16. Bryson G. When is a PAS required? What is the utility of a PAS? Presented at: Greenberg Stent Summit XVI; December 4-6, 2023; Boston, Massachusetts.
17. Eagleton MJ, Follansbee M, Wolski K, et al. Fenestrated and branched endovascular aneurysm repair outcomes for type II and III thoracoabdominal aortic aneurysms. *J Vasc Surg.* 2016;63:930-942. doi: 10.1016/j.jvs.2015.10.095

(Continued on page 55)

(Continued from page 53)

18. Parodi EF, Schanzer A, Oderich GS, et al. The development and potential implications of the US Fenestrated and Branched Aortic Research Consortium. *Semin Vasc Surg.* 2022;35:380-384. doi: 10.1053/j.semvasc-surg.2022.08.002

19. Eid MA, Barry MJ, Tang GL, et al. Effect of a decision aid on agreement between patient preferences and repair

type for abdominal aortic aneurysm: a randomized clinical trial. *JAMA Surg.* 2022;157:e222935. doi:10.1001/jamasurg.2022.2935

20. US Food and Drug Administration. Patient preference information – voluntary submission, review in premarket approval applications, humanitarian device exemption applications, and de novo requests, and inclusion in decision summaries and device labeling. Guidance for industry, Food and Drug Administration staff, and other stakeholders. Accessed February 15, 2024. <https://www.fda.gov/media/92593/download>

Rohini Retarekar, PhD

Team Lead, Vascular and Endovascular Devices Team (VEDT)

Division of Circulatory Support, Structural and Vascular Devices

Office of Cardiovascular Devices

Office of Product Evaluation and Quality

CDRH, US Food and Drug Administration

Silver Spring, Maryland

Disclosures: None.

Robert Lee, MD

Medical Officer, Vascular and Endovascular Devices Team (VEDT)

Division of Circulatory Support, Structural and Vascular Devices

Office of Cardiovascular Devices

CDRH, US Food and Drug Administration

Silver Spring, Maryland

Disclosures: None.

Ronald Fairman, MD

Medical Officer, Vascular and Endovascular Devices Team (VEDT)

Division of Circulatory Support, Structural and Vascular Devices

Office of Cardiovascular Devices

Office of Product Evaluation and Quality

CDRH, US Food and Drug Administration

Silver Spring, Maryland

Disclosures: None.

Carmen Gacchina Johnson, PhD

Assistant Director, Vascular and Endovascular Devices Team (VEDT)

Division of Circulatory Support, Structural and Vascular Devices

Office of Cardiovascular Devices

Office of Product Evaluation and Quality

CDRH, US Food and Drug Administration

Silver Spring, Maryland

carmen.gacchina@fda.hhs.gov

Disclosures: None.