

Sublime™ .014, .018, and .035 Microcatheters

INDICATIONS FOR USE

The Sublime Microcatheter is intended to access the peripheral vasculature in order to facilitate the placement and/or the exchange of guidewires. The Sublime Microcatheter is also intended to provide a conduit for the delivery of saline solutions or diagnostic contrast agents.

CAUTION: Federal (US) law restricts these devices to sale by or on the order of a physician. Please refer to Instructions for Use for indications, contraindications, warnings, and precautions.

CONTRAINDICATIONS

None known.

Sublime™ Radial Access Guide Sheath

INDICATIONS FOR USE

The Sublime Guide Sheath is intended to introduce therapeutic or diagnostic devices into the vasculature, excluding the coronary and neurovasculature.

CAUTION: Federal (US) law restricts these devices to sale by or on the order of a physician. Please refer to Instructions for Use for indications, contraindications, warnings, and precautions.

Sublime™ Radial Access 014 RX PTA Dilatation Catheter

INDICATIONS FOR USE

The Sublime™ Radial Access 014 RX PTA Dilatation Catheter is indicated for Percutaneous Transluminal Angioplasty (PTA) dilation of peripheral vasculature stenoses in the iliac, femoral, ilio-femoral, popliteal, infrapopliteal, and renal arteries, and for the treatment of obstructive lesions of native or synthetic arteriovenous dialysis fistulae.

CAUTION: Federal (US) law restricts these devices to sale by or on the order of a physician. Please refer to Instructions for Use for indications, contraindications, warnings, and precautions.

CONTRAINDICATIONS

No contraindications known for PTA procedure. The 014 Rx PTA Balloon Catheter is contraindicated for use in the coronary arteries and the neurovasculature.

Sublime™ Radial Access 018 RX PTA Dilatation Catheter

INDICATIONS FOR USE

The Sublime™ Radial Access 018 RX PTA Dilatation Catheter is indicated for Percutaneous Transluminal Angioplasty (PTA) dilation of peripheral vasculature stenoses in the iliac, femoral, ilio-femoral, popliteal, infrapopliteal, and renal arteries, and for the treatment of obstructive lesions of native or synthetic arteriovenous dialysis fistulae.

CAUTION: Federal (US) law restricts these devices to sale by or on the order of a physician. Please refer to Instructions for Use for indications, contraindications, warnings, and precautions.

CONTRAINDICATIONS

No contraindications known for PTA procedure. The 018 Rx PTA Balloon Catheter is contraindicated for use in the coronary arteries and the neurovasculature.

The opinions, clinical and otherwise, presented here are informational only. The opinions are those of the presenter only and do not necessarily reflect the views of Surmodics. Results discussed from use of Surmodics or other products may not be predictive of all patients and may vary depending on differing patient characteristics.